

# Improving Reporting of High Radiation Dose Fluoroscopy Events to Meet New Joint Commission Requirements

Phuong-Anh T. Duong MD<sup>1</sup>, Shalmali Dharmadhikari PhD<sup>2</sup>, Rebecca Neill MS<sup>2</sup>, Matthew Zygmunt MD<sup>3</sup>

<sup>1</sup>Department of Radiology and Imaging Sciences, University of Utah <sup>2</sup>Environmental Health and Safety Office, Emory University, <sup>3</sup>Department of Radiology and Imaging Sciences, Emory University




## Aim Statement

The goal is to improve the detection rate of high radiation dose fluoroscopy events by using a semi-automated reporting process.

## Background

- Complex procedures in interventional radiology often result in high fluoroscopic doses.
- Patients receiving high doses of radiation require monitoring to assess for skin injury.
- As of January 2019, the Joint Commission mandates recording fluoroscopy dose information, establishing radiation dose investigation thresholds, and following patients when thresholds are exceeded.
- Our original process required the operator to self report high doses and was unreliable. For example, some events would go unreported until the patient is seen in clinic.

## Analysis



## Actions/Tests of Change

Test of Change	Date	Result
Added capable equipment to dose tracking system which generates alerts for high doses	5/2016-12/2017	Increased number of cases detected.
Added <b>cumulative</b> dose alert 6 month dose >15 Gy	12/2017-	No events during this time period
Reviewed monthly reports of required manually recorded CAK in (RIS)	1/2019 -	Discovered 7 high dose events that were missed due to sending errors or older equipment


## Results

- Dose tracking challenging to implement—some equipment required software upgrades or working with manufacturer to enable data transfer.
- Dose tracking system detected dose trends excessive dose trends resulting in the discovery of a malfunctioning unit which was decommissioned until repaired.
- Increased number of events detected; more so far in first 3 quarters this year than previous years.
- RIS audit revealed 7 cases missed by dose tracking software due to sending errors and from older equipment that could not send data.

## Baseline Conditions

- Interventional staff were expected to note cumulative air kerma (CAK) for each case in the radiology information system (RIS).
- If CAK exceeded a threshold of 5 Gy, a skin injury dose risk estimate form was completed and faxed to the Environmental Health and Safety Office (EHSO).
- Peak skin dose (PSD) was calculated by a medical physicist using additional information about the acquisitions from the fluoroscopy unit.
- Due to periodic deletion, delays in reporting would result in loss of information needed to calculate skin doses.
- Cumulative patient doses were difficult to track when using manual tracking of doses.



## Baseline number of high dose events over time



## Measures

Number of high radiation dose cases detected by faxed forms, dose tracking software, and monthly audit of manual entered dose in the RIS.

## Revised Process (new steps in blue)



## Reflection/Follow-up

- Dose tracking system revealed underreporting of excessive dose events.
- Audit of manual entry in RIS was a remediation step to improve reporting reliability.
- Automated alerts enabled us to detect trends which revealed malfunctioning equipment earlier.
- Automated recording of doses also allowed tracking of cumulative patient exposure.
- Dose tracking planned for fluoroscopy units in other procedure areas (Cardiology) which continue to rely on faxed forms.